

## Test Report

1. Model : AL-5000

2. Description : Breath Alcohol Tester

3. Test Date : 2004.9.07

4. Environment of Labartory

- . Temperature :  $23\pm 1.5$  °C
- . Relative Humidity :  $50\pm 10\%$
- . Simulator Test Pressure : 20 λ /min

5. .Test simulator :

Guth Laboratories Inc. Guth Model 34C , Serial G 6211)

95% saturated in water at 34 celsius degree conditions.

- 1) Test concentration 0,40 g/l in blood
- 2) Test concentration 0,60 g/l in blood
- 3) Calibration concentration 0,50 g/l and 1.00g/l in blood

6. Display unit

- . g/l

7. Test result

	Sample	#1	Sample	#2	Sample	#3
	0.4g/l	0.6g/l	0.4g/l	0.6g/l	0.4g/l	0.6g/l
#1	0.4	0.5	0.4	0.6	0.4	0.6
#2	0.4	0.6	0.4	0.6	0.4	0.6
#3	0.4	0.6	0.4	0.6	0.4	0.5
#4	0.4	0.6	0.4	0.6	0.4	0.5
#5	0.4	0.6	0.4	0.6	0.4	0.6
#6	0.4	0.6	0.3	0.6	0.4	0.6
#7	0.4	0.6	0.4	0.6	0.4	0.6
#8	0.4	0.6	0.4	0.6	0.4	0.6
#9	0.4	0.6	0.4	0.6	0.4	0.6
#10	0.4	0.6	0.4	0.6	0.4	0.6
#11	0.4	0.6	0.4	0.6	0.4	0.6
#12	0.4	0.6	0.4	0.6	0.4	0.6
#13	0.4	0.6	0.4	0.6	0.4	0.6
#14	0.4	0.6	0.4	0.6	0.4	0.6
#15	0.4	0.6	0.4	0.6	0.4	0.6
#16	0.4	0.6	0.4	0.6	0.4	0.6
#17	0.4	0.6	0.4	0.6	0.4	0.6
#18	0.4	0.6	0.4	0.6	0.4	0.6
#19	0.4	0.6	0.5	0.6	0.4	0.6
#20	0.4	0.6	0.4	0.6	0.4	0.6
#21	0.4	0.6	0.4	0.6	0.4	0.6
#22	0.4	0.6	0.4	0.6	0.4	0.6
#23	0.4	0.6	0.4	0.6	0.4	0.6
#24	0.4	0.6	0.4	0.7	0.4	0.6
#25	0.4	0.6	0.4	0.6	0.4	0.6
#26	0.4	0.6	0.4	0.6	0.4	0.5
#27	0.4	0.6	0.4	0.6	0.5	0.6

#28	0.4	0.6	0.4	0.6	0.4	0.6
#29	0.4	0.6	0.4	0.6	0.4	0.6
#30	0.4	0.6	0.4	0.6	0.4	0.6
#31	0.4	0.6	0.4	0.6	0.4	0.6
#32	0.4	0.6	0.4	0.6	0.4	0.6
#33	0.4	0.6	0.4	0.6	0.4	0.6
#34	0.4	0.6	0.4	0.6	0.4	0.6
#35	0.4	0.6	0.4	0.6	0.4	0.5
#36	0.4	0.6	0.4	0.6	0.4	0.6
#37	0.4	0.6	0.4	0.6	0.4	0.6
#38	0.4	0.6	0.4	0.6	0.4	0.6
#39	0.4	0.6	0.4	0.6	0.4	0.6
#40	0.4	0.6	0.4	0.6	0.4	0.6
#41	0.4	0.6	0.4	0.6	0.4	0.6
#42	0.4	0.6	0.4	0.6	0.4	0.6
#43	0.4	0.6	0.4	0.6	0.4	0.6
#44	0.4	0.6	0.4	0.6	0.4	0.6
#45	0.4	0.6	0.4	0.6	0.4	0.6
#46	0.4	0.6	0.4	0.6	0.4	0.6
#47	0.4	0.6	0.4	0.6	0.4	0.6
#48	0.4	0.6	0.4	0.6	0.4	0.6
#49	0.4	0.6	0.4	0.6	0.4	0.6
#50	0.4	0.6	0.4	0.6	0.5	0.6

---